



金华诺科生物科技有限公司
Jinhua Noke Biotechnology Co., Ltd
产品质检单
Certificate of Analysis

No.: NK/R-B-017/6

产品名称 Commodity	离心管 Centrifuge tube	生产批号 Batch NO.	20250613	
编号 Part Number	NKCT-15-C-N	生产日期 Manufacture Date	2025.06.13	
有效期 Valid Date	3年 (Years)	报告日期 Report date	2025年6月14日	
规格型号 Size	15ml, 50支/包, 非灭菌, PP材质无重金属 15ml, 50pcs/pack, non sterile, PP material without heavy metalls	数量 Quantity	2000	支 pcs
生产依据 ISO 9001 Quality Standard	该产品是在质量管理体系下制造的, 并获得了ISO 9001质量体系标准认证的注册机构的批准。 The product was manufactured under Quality Management System and is approved by an accredited registering body to the ISO 9001 Quality Systems Standard.			
测试方法和依据 Testing Method	企业标准 Enterprise standards			
检测项目 Testing Item	检测方式 Detection method	企业标准 Enterprise Standards		检测结果 Test results
外观 Appearance	目测/污点比对卡 Visual inspection/stain comparison card	1. 无明显拉伤、划痕、裂痕; 无明显气泡、无开裂、无碎屑、无异物等; 1.no obvious strains, scratches, cracks; no obvious bubbles, no cracks, no debris, no foreign objects, etc.; 2. 产品洁净, 无脏污, 无收缩, 无油污, 无黑点; 2.clean, no dirt, shrinkage, oil stains or black spots 3. 印刷清晰、完整; 3. Printing is clear and complete.		合格 Qualified
包装及包装材料 Packaging & Packaging Materials	目测 Visual inspection	内包装密封不漏气, 外包装无明显的脏污, 无破损, 无变形等现象; 标识、标签清晰, 标签内容和箱内的产品相符。 Inner packaging is airtight; outer packaging is free from visible dirt, damage, and deformation. Labels are clear, and label content matches the products inside the box.		合格 Qualified
尺寸 Sizes	量器测量 Gauge	符合产品图纸设计要求。 Meets the design requirements of the product drawings		合格 Qualified
离心管密封性 Tube Sealness	密封性测试仪 Sealing tester	抽负压 -90Kpa, 30分钟无渗漏 Draw negative pressure -90Kpa, no leakage for 30 minutes		合格 Qualified
离心力 Tube centrifugal force	离心机 Centrifuge	符合企业标准, 模具设计标准 Conform to Company standard		合格 Qualified
印刷效果 Printing effect	目测 Visual inspection	1. 引述内容正确清晰, 无缺墨、重影现象 1.The printed content is correct and clear, without ink shortage or ghosting 2. 印刷牢固, 3M胶带粘贴3次无脱墨 2. Printing is firm, 3M tape is pasted 3 times without deinking		合格 Qualified
灭菌 Sterilization	目测 Visual inspection	EO灭菌, 有《EO灭菌证明》。 Has a Certificate of EO sterile.		
RESULT: 综合判别:	QUALIFIED (√) 合格 (√)	UNQUALIFIED () 不合格 ()		
SUGGESTION: 建议:	RELEASE (√) 准允出库 (√)	REWORK () 返工 ()	SCRAP () 报废 ()	
CHECKED BY: PENG TING 质检人: PENG TING	DATE: 14/06/2025 日期: 2025/6/14	APPROVED BY: CHEN 审核人: Chen	DATE: 14/6/2025 日期: 2025/06/14	

