



金华诺科生物科技有限公司
Jinhua Noke Biotechnology Co., Ltd
产品质检单

Certificate of Analysis

产品名称 Commodity	吸头 Pipette tips	生产批号 Batch NO.	26080301	
编号 Part Number	NKPT-5K-E-B (new number: NKPT-5K-E-B1)	生产日期 Manufacture Date	2026.08.03	
有效期 Valid Date	5 年 (Years)	报告日期 Report date	2026年8月15日	
规格型号 Size	5ml 大口吸头, 适配艾本德, PP料无重金属 5ml pipette tips, bigger mouth, EP tyep, PP material without any metals	数量 Quantity	300	包 bag
生产依据 ISO 9001 Quality Standard	该产品是在质量管理体系下制造的, 并获得了ISO 9001质量体系标准认证的注册机构的批准。 The product was manufactured under Quality Management System and is approved by an accredited registering body to the ISO 9001 Quality Systems Standard.			
测试方法和依据 Testing Method	企业标准 Enterprise standards			
检测项目 Testing Item	检测方式 Detection method	质量要求 Requirement	检测结果 Test results	
外观 Appearance	目测/污点比对卡 Visual inspection/stain comparison card	产品表面应光滑平整透明, , 无毛刺、气泡、缺损、划痕等缺陷。 The surface of the product should be smooth, flat, transparent, and free of defects such as burrs, bubbles, defects, scratches, etc.	合格 Qualified	
包装及包装材料 Packaging & Packaging Materials	目测 Visual inspection	内包装密封不漏气, 外包装无明显的脏污,无破损,无变形等现象; 标识、标签清晰, 标签内容和箱内的产品相符。 Inner packaging is airtight; outer packaging is free from visible dirt, damage, and deformation. Labels are clear, and label content matches the products inside the box.	合格 Qualified	
尺寸 Sizes	量器测量 Gauge	符合产品图纸设计要求。 Meets the design requirements of the product drawings	合格 Qualified	
适配性 adaptability	移液器 Micropipette	移液枪抓取吸头并吸液5ml,检测1分钟不漏液。 The pipette captures the pipette tip and sucks 5ml of liquid, testing for 1 minute without any leakage.	合格 Qualified	
垂直度 verticality	影像测量仪 Image measuring instrument	≤1.5mm	合格 Qualified	
灭菌 Sterilization	目测 Visual inspection	辐照标签呈红色, 有《辐照证明》。 The irradiation label is red in color and has a Certificate of Irradiation.	/	
残留值 Residual	移液器 Micropipette	吸头吸取测试液体, 净置10s后完全排除液体, 通过称重法计算残留值, 残留值<1%。 Suck liquid with tip, sit for 10 seconds, and completely remove the liquid. Weigh the residual value, should be less than 1%.	合格 Qualified	
内毒素检测 Endotoxin testing	《细菌内毒素检查规范》 "Bacterial Endotoxin Inspection Specification"	符合《细菌内毒素检查规范》 Confirm to "Bacterial Endotoxin Inspection Specification"	/	
核酸酶检测 Rnase detection	《核酸酶测试规范》 "Nuclease Test Specification"	符合《核酸酶测试规范》 Confirm to "Nuclease Test Specification"	/	
DNA酶检测 Dnase detection	《核酸酶测试规范》 "Nuclease Test Specification"	符合《核酸酶测试规范》 Confirm to "Nuclease Test Specification"	/	
RESULT: 综合判别:	QUALIFIED (√) 合格 (√)	UNQUALIFIED () 不合格 ()		
SUGGESTION: 建议:	RELEASE (√) 准允出库 (√)	REWORK () 返工 ()	SCRAP () 报废 ()	
CHECKED BY: PENG TING 质检人: PENG TING	DATE: 15/08/2026 日期: 15/08/2026	APPROVED BY: CHEN 审核人: Chen	DATE: 15/08/2026 日期: 15/08/2026	质量部 检验专用章